

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
 Data File : VN054924.D
 Acq On : 3 Apr 2019 14:15
 Operator : JC/SP
 Sample : K2214-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2-S

Quant Time: Apr 04 02:19:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	496425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	837065	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	727856	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	276053	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	325090	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
35) Dibromofluoromethane	7.59	113	264069	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	10.09	98	1005139	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
62) 4-Bromofluorobenzene	12.40	95	321451	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
Target Compounds						
					Qvalue	
16) Acetone	3.82	43	15884	9.954	ug/l #	86
20) Methylene Chloride	4.55	84	238017	42.284	ug/l #	83
43) Isopropyl Acetate	8.17	43	16564	1.782	ug/l #	93
67) Ethyl Benzene	11.51	91	33938	1.320	ug/l	94
68) m/p-Xylenes	11.62	106	21787	2.289	ug/l	96
78) n-propylbenzene	12.59	91	36868	1.487	ug/l	96
80) 1,3,5-Trimethylbenzene	12.73	105	51292	2.770	ug/l	94
84) 1,2,4-Trimethylbenzene	13.04	105	140556	7.969	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040319\
Data File : VN054924.D
Acq On : 3 Apr 2019 14:15
Operator : JC/SP
Sample : K2214-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2-S

Quant Time: Apr 04 02:19:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

