

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040319\
 Data File : VN054932.D
 Acq On : 3 Apr 2019 18:39
 Operator : JC/SP
 Sample : K1696-03 100X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 21860

Quant Time: Apr 04 03:22:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	503466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	838820	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	711357	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	251599	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	329656	48.57	ug/l	0.00
Spiked Amount			Recovery	=	97.14%	
35) Dibromofluoromethane	7.59	113	273448	48.90	ug/l	0.00
Spiked Amount			Recovery	=	97.80%	
50) Toluene-d8	10.09	98	1006810	48.43	ug/l	0.00
Spiked Amount			Recovery	=	96.86%	
62) 4-Bromofluorobenzene	12.40	95	304510	44.04	ug/l	0.00
Spiked Amount			Recovery	=	88.08%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040319\
Data File : VN054932.D
Acq On : 3 Apr 2019 18:39
Operator : JC/SP
Sample : K1696-03 100X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
21860

Quant Time: Apr 04 03:22:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

