


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN040319\
Data File : VN054913.D
Acq On : 3 Apr 2019 8:42
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Title : SW846 8260
Last Update : Sat Mar 30 00:48:18 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	12535	PASS
75	95	30	60	52.9	31338	PASS
95	95	100	100	100.0	59296	PASS
96	95	5	9	7.1	4216	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	83.1	49298	PASS
175	174	5	9	7.3	3612	PASS
176	174	95	101	95.8	47240	PASS
177	176	5	9	7.2	3385	PASS