

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
 Data File : VN054950.D
 Acq On : 4 Apr 2019 16:26
 Operator : JC/SP
 Sample : K2258-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STONE

Quant Time: Apr 05 03:46:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 05 02:23:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	542181	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	872028	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	762579	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	280936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	348152	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
35) Dibromofluoromethane	7.59	113	280014	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
50) Toluene-d8	10.09	98	1053609	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
62) 4-Bromofluorobenzene	12.40	95	329904	45.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.80%	
Target Compounds						
20) Methylene Chloride	4.55	84	215704	32.512	ug/l	99
43) Isopropyl Acetate	8.17	43	16508	1.652	ug/l #	93
95) Naphthalene	15.13	128	27411	5.298	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
Data File : VN054950.D
Acq On : 4 Apr 2019 16:26
Operator : JC/SP
Sample : K2258-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STONE

Quant Time: Apr 05 03:46:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 05 02:23:47 2019
Response via : Initial Calibration

