

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
 Data File : VN054952.D
 Acq On : 4 Apr 2019 17:18
 Operator : JC/SP
 Sample : K2245-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3S

Quant Time: Apr 05 03:52:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 05 02:23:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	520086	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871512	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	736315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	269611	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	342679	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
35) Dibromofluoromethane	7.59	113	281066	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
50) Toluene-d8	10.09	98	1042384	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
62) 4-Bromofluorobenzene	12.40	95	315355	43.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.80%	
Target Compounds						
20) Methylene Chloride	4.55	84	17221	2.706	ug/l	92
43) Isopropyl Acetate	8.17	43	16791	1.681	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
Data File : VN054952.D
Acq On : 4 Apr 2019 17:18
Operator : JC/SP
Sample : K2245-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3S

Quant Time: Apr 05 03:52:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 05 02:23:47 2019
Response via : Initial Calibration

