

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
 Data File : VN054954.D
 Acq On : 4 Apr 2019 18:11
 Operator : JC/SP
 Sample : K2245-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1S

Quant Time: Apr 05 03:55:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 05 02:23:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	514890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	866797	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	735923	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	271027	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	341124	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	7.59	113	275727	48.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.74%	
50) Toluene-d8	10.09	98	1027143	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
62) 4-Bromofluorobenzene	12.40	95	317966	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
Target Compounds						
20) Methylene Chloride	4.55	84	26268	4.169	ug/l	91
43) Isopropyl Acetate	8.17	43	17522	1.764	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
Data File : VN054954.D
Acq On : 4 Apr 2019 18:11
Operator : JC/SP
Sample : K2245-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1S

Quant Time: Apr 05 03:55:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 05 02:23:47 2019
Response via : Initial Calibration

