

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
 Data File : VN054958.D
 Acq On : 4 Apr 2019 19:57
 Operator : JC/SP
 Sample : K2193-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 HR-06-040319-A

Quant Time: Apr 05 04:07:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 05 02:23:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	520477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	872557	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	760598	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	306386	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	343972	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
35) Dibromofluoromethane	7.59	113	275204	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	10.09	98	1048907	48.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.34%	
62) 4-Bromofluorobenzene	12.40	95	340324	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
Target Compounds						
20) Methylene Chloride	4.55	84	565961	88.862	ug/l	98
43) Isopropyl Acetate	8.17	43	16817	1.682	ug/l	96
80) 1,3,5-Trimethylbenzene	12.73	105	48583	2.286	ug/l	97
84) 1,2,4-Trimethylbenzene	13.04	105	34498	1.677	ug/l	97
95) Naphthalene	15.13	128	20923	4.569	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN040419\
Data File : VN054958.D
Acq On : 4 Apr 2019 19:57
Operator : JC/SP
Sample : K2193-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
HR-06-040319-A

Quant Time: Apr 05 04:07:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 05 02:23:47 2019
Response via : Initial Calibration

