

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_N\DATA\VN040419\  
 Data File : VN054959.D  
 Acq On : 4 Apr 2019 20:23  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Apr 05 04:09:29 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N040419W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Apr 05 02:23:47 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 1 I   | Pentafluorobenzene          | 1.000 | 1.000 | 0.0   | 97    | 0.00     |
| 2 T   | Dichlorodifluoromethane     | 0.598 | 0.613 | -2.5  | 96    | 0.00     |
| 3 P   | Chloromethane               | 0.790 | 0.781 | 1.1   | 95    | 0.00     |
| 4 C   | Vinyl Chloride              | 0.691 | 0.686 | 0.7#  | 95    | 0.00     |
| 5 T   | Bromomethane                | 0.394 | 0.371 | 5.8   | 92    | 0.00     |
| 6 T   | Chloroethane                | 0.385 | 0.370 | 3.9   | 95    | 0.00     |
| 7 T   | Trichlorofluoromethane      | 0.839 | 0.836 | 0.4   | 96    | 0.00     |
| 8 T   | Diethyl Ether               | 0.267 | 0.266 | 0.4   | 97    | 0.00     |
| 9 T   | 1,1,2-Trichlorotrifluoroeth | 0.444 | 0.428 | 3.6   | 94    | 0.00     |
| 10 T  | Methyl Iodide               | 0.626 | 0.634 | -1.3  | 94    | 0.00     |
| 11 T  | Tert butyl alcohol          | 0.034 | 0.041 | -20.6 | 119   | 0.00     |
| 12 CM | 1,1-Dichloroethene          | 0.441 | 0.429 | 2.7#  | 95    | 0.00     |
| 13 T  | Acrolein                    | 0.047 | 0.057 | -21.3 | 106   | 0.00     |
| 14 T  | Allyl chloride              | 0.978 | 1.029 | -5.2  | 97    | 0.00     |
| 15 T  | Acrylonitrile               | 0.204 | 0.223 | -9.3  | 105   | 0.00     |
| 16 T  | Acetone                     | 0.189 | 0.151 | 20.1  | 78    | 0.00     |
| 17 T  | Carbon Disulfide            | 1.403 | 1.396 | 0.5   | 94    | 0.00     |
| 18 T  | Methyl Acetate              | 0.491 | 0.482 | 1.8   | 106   | 0.00     |
| 19 T  | Methyl tert-butyl Ether     | 1.430 | 1.525 | -6.6  | 102   | 0.00     |
| 20 T  | Methylene Chloride          | 0.612 | 0.609 | 0.5   | 100   | 0.00     |
| 21 T  | trans-1,2-Dichloroethene    | 0.550 | 0.558 | -1.5  | 96    | 0.00     |
| 22 T  | Diisopropyl ether           | 2.009 | 2.117 | -5.4  | 98    | 0.00     |
| 23 T  | Vinyl Acetate               | 1.398 | 1.583 | -13.2 | 103   | 0.00     |
| 24 P  | 1,1-Dichloroethane          | 1.096 | 1.118 | -2.0  | 98    | 0.00     |
| 25 T  | 2-Butanone                  | 0.248 | 0.274 | -10.5 | 99    | 0.00     |
| 26 T  | 2,2-Dichloropropane         | 0.768 | 0.716 | 6.8   | 90    | 0.00     |
| 27 T  | cis-1,2-Dichloroethene      | 0.624 | 0.639 | -2.4  | 98    | 0.00     |
| 28 T  | Bromochloromethane          | 0.513 | 0.527 | -2.7  | 100   | 0.00     |
| 29 T  | Tetrahydrofuran             | 0.162 | 0.187 | -15.4 | 109   | 0.00     |
| 30 C  | Chloroform                  | 1.038 | 1.061 | -2.2# | 98    | 0.00     |
| 31 T  | Cyclohexane                 | 1.251 | 1.052 | 15.9  | 96    | 0.00     |
| 32 T  | 1,1,1-Trichloroethane       | 0.876 | 0.883 | -0.8  | 97    | 0.00     |
| 33 S  | 1,2-Dichloroethane-d4       | 0.659 | 0.717 | -8.8  | 103   | 0.00     |
| 34 I  | 1,4-Difluorobenzene         | 1.000 | 1.000 | 0.0   | 100   | 0.00     |
| 35 S  | Dibromofluoromethane        | 0.325 | 0.335 | -3.1  | 100   | 0.00     |
| 36 T  | 1,1-Dichloropropene         | 0.495 | 0.483 | 2.4   | 96    | 0.00     |
| 37 T  | Ethyl Acetate               | 0.340 | 0.375 | -10.3 | 107   | 0.00     |
| 38 T  | Carbon Tetrachloride        | 0.478 | 0.457 | 4.4   | 96    | 0.00     |
| 39 T  | Methylcyclohexane           | 0.545 | 0.538 | 1.3   | 95    | 0.00     |
| 40 TM | Benzene                     | 1.426 | 1.436 | -0.7  | 97    | 0.00     |
| 41 T  | Methacrylonitrile           | 0.194 | 0.223 | -14.9 | 112   | 0.00     |
| 42 TM | 1,2-Dichloroethane          | 0.501 | 0.503 | -0.4  | 99    | 0.00     |
| 43 T  | Isopropyl Acetate           | 0.573 | 0.649 | -13.3 | 109   | 0.00     |
| 44 TM | Trichloroethene             | 0.365 | 0.368 | -0.8  | 99    | 0.00     |
| 45 C  | 1,2-Dichloropropane         | 0.397 | 0.403 | -1.5# | 99    | 0.00     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_N\DATA\VN040419\  
 Data File : VN054959.D  
 Acq On : 4 Apr 2019 20:23  
 Operator : JC/SP  
 Sample : VSTDCCC050  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 LabSampleId :  
 VSTDCCC050

Quant Time: Apr 05 04:09:29 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N040419W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Apr 05 02:23:47 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|       | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|-------|-----------------------------|-------|-------|-------|-------|----------|
| 46 T  | Dibromomethane              | 0.225 | 0.230 | -2.2  | 100   | 0.00     |
| 47 T  | Bromodichloromethane        | 0.473 | 0.484 | -2.3  | 99    | 0.00     |
| 48 T  | Methyl methacrylate         | 0.290 | 0.325 | -12.1 | 106   | 0.00     |
| 49 T  | 1,4-Dioxane                 | 0.003 | 0.003 | 0.0   | 111   | 0.00     |
| 50 S  | Toluene-d8                  | 1.235 | 1.267 | -2.6  | 100   | 0.00     |
| 51 T  | 4-Methyl-2-Pentanone        | 0.320 | 0.359 | -12.2 | 107   | 0.00     |
| 52 CM | Toluene                     | 0.840 | 0.854 | -1.7# | 96    | 0.00     |
| 53 T  | t-1,3-Dichloropropene       | 0.437 | 0.474 | -8.5  | 100   | 0.00     |
| 54 T  | cis-1,3-Dichloropropene     | 0.532 | 0.559 | -5.1  | 98    | 0.00     |
| 55 T  | 1,1,2-Trichloroethane       | 0.305 | 0.313 | -2.6  | 99    | 0.00     |
| 56 T  | Ethyl methacrylate          | 0.398 | 0.441 | -10.8 | 103   | 0.00     |
| 57 T  | 1,3-Dichloropropane         | 0.532 | 0.554 | -4.1  | 100   | 0.00     |
| 58 T  | 2-Chloroethyl Vinyl ether   | 0.157 | 0.177 | -12.7 | 104   | 0.00     |
| 59 T  | 2-Hexanone                  | 0.200 | 0.226 | -13.0 | 104   | 0.00     |
| 60 T  | Dibromochloromethane        | 0.319 | 0.338 | -6.0  | 99    | 0.00     |
| 61 T  | 1,2-Dibromoethane           | 0.286 | 0.311 | -8.7  | 102   | 0.00     |
| 62 S  | 4-Bromofluorobenzene        | 0.412 | 0.431 | -4.6  | 101   | 0.00     |
| 63 I  | Chlorobenzene-d5            | 1.000 | 1.000 | 0.0   | 98    | 0.00     |
| 64 T  | Tetrachloroethene           | 0.403 | 0.392 | 2.7   | 96    | 0.00     |
| 65 PM | Chlorobenzene               | 1.019 | 1.021 | -0.2  | 96    | 0.00     |
| 66 T  | 1,1,1,2-Tetrachloroethane   | 0.373 | 0.390 | -4.6  | 99    | 0.00     |
| 67 C  | Ethyl Benzene               | 1.838 | 1.861 | -1.3# | 96    | 0.00     |
| 68 T  | m/p-Xylenes                 | 0.672 | 0.686 | -2.1  | 96    | 0.00     |
| 69 T  | o-Xylene                    | 0.655 | 0.674 | -2.9  | 97    | 0.00     |
| 70 T  | Styrene                     | 1.051 | 1.122 | -6.8  | 96    | 0.00     |
| 71 P  | Bromoform                   | 0.224 | 0.241 | -7.6  | 100   | 0.00     |
| 72 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0   | 98    | 0.00     |
| 73 T  | Isopropylbenzene            | 4.224 | 3.965 | 6.1   | 95    | 0.00     |
| 74 T  | N-amyl acetate              | 1.174 | 1.354 | -15.3 | 106   | 0.00     |
| 75 P  | 1,1,2,2-Tetrachloroethane   | 1.037 | 1.017 | 1.9   | 102   | 0.00     |
| 76 T  | 1,2,3-Trichloropropane      | 0.899 | 0.843 | 6.2   | 96    | 0.00     |
| 77 T  | Bromobenzene                | 1.023 | 0.988 | 3.4   | 97    | 0.00     |
| 78 T  | n-propylbenzene             | 4.653 | 4.482 | 3.7   | 95    | 0.00     |
| 79 T  | 2-Chlorotoluene             | 2.884 | 2.722 | 5.6   | 96    | 0.00     |
| 80 T  | 1,3,5-Trimethylbenzene      | 3.469 | 3.340 | 3.7   | 96    | 0.00     |
| 81 T  | trans-1,4-Dichloro-2-butene | 0.216 | 0.237 | -9.7  | 105   | 0.00     |
| 82 T  | 4-Chlorotoluene             | 2.728 | 2.693 | 1.3   | 97    | 0.00     |
| 83 T  | tert-Butylbenzene           | 2.951 | 2.776 | 5.9   | 95    | 0.00     |
| 84 T  | 1,2,4-Trimethylbenzene      | 3.356 | 3.283 | 2.2   | 95    | 0.00     |
| 85 T  | sec-Butylbenzene            | 3.817 | 3.642 | 4.6   | 96    | 0.00     |
| 86 T  | p-Isopropyltoluene          | 3.289 | 3.195 | 2.9   | 95    | 0.00     |
| 87 T  | 1,3-Dichlorobenzene         | 1.702 | 1.665 | 2.2   | 96    | 0.00     |
| 88 T  | 1,4-Dichlorobenzene         | 1.620 | 1.621 | -0.1  | 97    | 0.00     |
| 89 T  | n-Butylbenzene              | 2.628 | 2.664 | -1.4  | 94    | 0.00     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_N\DATA\VN040419\  
Data File : VN054959.D  
Acq On : 4 Apr 2019 20:23  
Operator : JC/SP  
Sample : VSTDCCC050  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
LabSampleId :  
VSTDCCC050

Quant Time: Apr 05 04:09:29 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N040419W.M  
Quant Title : SW846 8260  
QLast Update : Fri Apr 05 02:23:47 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|-------|-------|-------|----------|
| 90 T | Hexachloroethane            | 0.552 | 0.552 | 0.0   | 97    | 0.00     |
| 91 T | 1,2-Dichlorobenzene         | 1.671 | 1.653 | 1.1   | 98    | 0.00     |
| 92 T | 1,2-Dibromo-3-Chloropropane | 0.145 | 0.154 | -6.2  | 107   | 0.00     |
| 93 T | 1,2,4-Trichlorobenzene      | 0.743 | 0.846 | -13.9 | 99    | 0.00     |
| 94 T | Hexachlorobutadiene         | 0.559 | 0.497 | 11.1  | 94    | 0.00     |
| 95 T | Naphthalene                 | 1.631 | 1.867 | -14.5 | 103   | 0.00     |
| 96 T | 1,2,3-Trichlorobenzene      | 0.767 | 0.842 | -9.8  | 101   | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 6