

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
 Data File : VN071822.D
 Acq On : 04 Apr 2022 10:51
 Operator : JC\MD
 Sample : VN0404MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0404MBL01

Quant Time: Apr 05 01:15:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	496550	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	844864	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	844144	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	294885	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	287615	50.737	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.480%
35) Dibromofluoromethane	8.022	113	243837	49.047	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.100%
50) Toluene-d8	10.439	98	1053532	51.803	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.600%
62) 4-Bromofluorobenzene	12.727	95	354513	53.898	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
Data File : VN071822.D
Acq On : 04 Apr 2022 10:51
Operator : JC\MD
Sample : VN0404MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0404MBL01

Quant Time: Apr 05 01:15:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

