

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
 Data File : VN071833.D
 Acq On : 04 Apr 2022 15:53
 Operator : JC\MD
 Sample : N2233-01RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6RE

Quant Time: Apr 05 01:18:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	631763	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1072835	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1099903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	416166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	354484	49.150	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.300%
35) Dibromofluoromethane	8.022	113	305583	48.405	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.820%
50) Toluene-d8	10.439	98	1345700	52.108	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.220%
62) 4-Bromofluorobenzene	12.727	95	471815	56.490	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	112.980%
Target Compounds						
16) Acetone	4.287	43	5322	1.727	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
Data File : VN071833.D
Acq On : 04 Apr 2022 15:53
Operator : JC\MD
Sample : N2233-01RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6RE

Quant Time: Apr 05 01:18:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

