

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
 Data File : VN071834.D
 Acq On : 04 Apr 2022 16:17
 Operator : JC\MD
 Sample : N2229-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AQUEOUS-DRUM

Quant Time: Apr 05 01:19:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	623131	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1058848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1071601	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	396626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	346437	48.700	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.400%
35) Dibromofluoromethane	8.016	113	301762	48.431	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.860%
50) Toluene-d8	10.439	98	1322161	51.873	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.740%
62) 4-Bromofluorobenzene	12.727	95	451087	54.721	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.440%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	8489	2.792	ug/l	99
20) Methylene Chloride	5.098	84	7592	1.185	ug/l	91
25) 2-Butanone	7.340	43	3430	0.783	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
Data File : VN071834.D
Acq On : 04 Apr 2022 16:17
Operator : JC\MD
Sample : N2229-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AQUEOUS-DRUM

Quant Time: Apr 05 01:19:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

