

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
 Data File : VN071843.D
 Acq On : 04 Apr 2022 19:53
 Operator : JC\MD
 Sample : N2219-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Apr 05 01:21:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	612123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1047839	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1087478	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	393403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	341384	48.852	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.700%
35) Dibromofluoromethane	8.022	113	301053	48.825	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.660%
50) Toluene-d8	10.439	98	1345406	53.340	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.680%
62) 4-Bromofluorobenzene	12.727	95	455466	55.833	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.660%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	58038	19.434	ug/l	99
20) Methylene Chloride	5.099	84	420317	66.773	ug/l	99
43) Isopropyl Acetate	8.557	43	45271	2.888	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040422\
Data File : VN071843.D
Acq On : 04 Apr 2022 19:53
Operator : JC\MD
Sample : N2219-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Apr 05 01:21:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

