


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN040519\
 Data File : VN054960.D
 Acq On : 5 Apr 2019 8:33
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N040419W.M
 Title : SW846 8260
 Last Update : Fri Apr 05 02:23:47 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.3	11285	PASS
75	95	30	60	51.5	28618	PASS
95	95	100	100	100.0	55570	PASS
96	95	5	9	6.9	3809	PASS
173	174	0.00	2	1.5	705	PASS
174	95	50	100	85.0	47210	PASS
175	174	5	9	6.7	3177	PASS
176	174	95	101	97.2	45885	PASS
177	176	5	9	6.1	2797	PASS