

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
 Data File : VN071859.D
 Acq On : 05 Apr 2022 16:24
 Operator : JC\MD
 Sample : N2258-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-E5-DUP

Quant Time: Apr 06 06:41:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	604502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1027979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1032666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	359414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	335192	48.571	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.140%
35) Dibromofluoromethane	8.022	113	295835	48.906	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.820%
50) Toluene-d8	10.439	98	1280497	51.747	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.500%
62) 4-Bromofluorobenzene	12.727	95	422071	52.739	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
Data File : VN071859.D
Acq On : 05 Apr 2022 16:24
Operator : JC\MD
Sample : N2258-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TW-E5-DUP

Quant Time: Apr 06 06:41:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

