

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
 Data File : VN071866.D
 Acq On : 05 Apr 2022 19:13
 Operator : JC\MD
 Sample : N2269-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LP-2

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/06/2022
 Supervised By : Mahesh Dadoda 04/06/2022

Quant Time: Apr 06 06:44:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	675741	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1152440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1172450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	454250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	376113	48.755	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.500%		
35) Dibromofluoromethane	8.022	113	328848	48.492	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.980%		
50) Toluene-d8	10.439	98	1431170	51.590	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.180%		
62) 4-Bromofluorobenzene	12.727	95	494074	55.069	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	110.140%		
Target Compounds						
16) Acetone	4.292	43	96356	29.227	ug/l #	88
25) 2-Butanone	7.345	43	9730m	2.047	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
Data File : VN071866.D
Acq On : 05 Apr 2022 19:13
Operator : JC\MD
Sample : N2269-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LP-2

Quant Time: Apr 06 06:44:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/06/2022
Supervised By :Mahesh Dadoda 04/06/2022

