

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
 Data File : VN071867.D
 Acq On : 05 Apr 2022 19:38
 Operator : JC\MD
 Sample : N2269-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Apr 06 06:44:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 04/06/2022
 Supervised By : Mahesh Dadoda 04/06/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	612603	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1049809	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1044189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	378108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	339874	48.598	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.200%		
35) Dibromofluoromethane	8.022	113	298433	48.309	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.620%		
50) Toluene-d8	10.439	98	1310189	51.846	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	103.700%		
62) 4-Bromofluorobenzene	12.727	95	429180	52.512	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	105.020%		
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	23121	7.736	ug/l	100
20) Methylene Chloride	5.099	84	6896	1.095	ug/l	90
25) 2-Butanone	7.340	43	5072m	1.177	ug/l	
30) Chloroform	7.810	83	21791	1.968	ug/l	94
47) Bromodichloromethane	9.763	83	9188	1.010	ug/l	99
60) Dibromochloromethane	11.233	129	3428	0.496	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040522\
Data File : VN071867.D
Acq On : 05 Apr 2022 19:38
Operator : JC\MD
Sample : N2269-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Apr 06 06:44:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/06/2022
Supervised By : Mahesh Dadoda 04/06/2022

