



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040523\
Data File : VN077253.D
Acq On : 05 Apr 2023 09:53
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040523W.M
Title : SW846 8260
Last Update : Thu Apr 06 08:27:03 2023



AutoFind: Scans 1854, 1855, 1856; Background Corrected with Scan 1845

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	22301	PASS
75	95	30	60	51.7	64421	PASS
95	95	100	100	100.0	124645	PASS
96	95	5	9	7.3	9084	PASS
173	174	0.00	2	0.5	462	PASS
174	95	50	100	73.4	91475	PASS
175	174	5	9	6.7	6147	PASS
176	174	95	101	98.2	89800	PASS
177	176	5	9	6.2	5534	PASS

Average of 12.842 to 12.854 min.: VN077253.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.90	1112	49.05	4611	59.95	1483	73.00	607
37.00	6374	50.00	22301	61.00	5759	74.00	2454
38.00	6584	51.00	7289	62.05	6500	75.00	6442
39.00	2602	52.05	322	63.00	5444	76.00	637
39.85	382	53.00	203	64.05	725	76.90	573
43.95	983	53.80	101	65.05	154	77.10	259
45.00	1053	54.95	571	67.10	199	77.70	111
45.70	67	56.00	2875	68.00	13638	77.90	520
46.10	206	56.95	4301	69.00	13485	78.95	5839
47.00	1621	57.85	252	70.05	1564	79.85	1466
47.95	941	58.90	70	72.05	1079	80.95	6011

Average of 12.842 to 12.854 min.: VN077253.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
81.80	457	94.00	16596	110.10	86	122.10	204
82.00	1065	95.00	124645	110.70	66	125.10	98
82.95	345	96.10	9084	110.95	243	127.80	213
84.30	76	96.95	549	111.95	176	128.05	307
85.05	290	102.90	107	112.95	222	128.80	294
85.75	155	103.95	780	114.85	288	129.90	559
86.95	5529	104.65	136	115.75	434	130.85	189
87.95	5697	105.05	280	116.95	923	131.10	90
90.95	795	105.85	765	117.90	599	134.85	241
91.95	3762	106.80	57	118.85	694	135.90	55
93.00	5812	109.80	122	121.10	110	136.80	243

Average of 12.842 to 12.854 min.: VN077253.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
137.10	129	152.10	61	174.00	91475		
140.10	71	154.10	64	175.00	6147		
140.85	1411	154.90	316	176.00	89800		
141.80	74	155.10	80	177.00	5534		
142.10	59	156.90	138	177.90	85		
142.95	1528	157.20	94	178.10	161		
145.80	167	170.70	56	206.85	24		
147.60	165	171.40	59	208.00	56		
147.90	82	171.80	61	281.40	56		
148.10	94	172.70	194				
149.90	55	172.95	462				