

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060858.D
 Acq On : 6 Apr 2020 17:50
 Operator : JC/MD
 Sample : L2144-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH

Quant Time: Apr 07 07:06:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336587	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	317462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149308	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	130305	46.85	ug/l	0.00
Spiked Amount			Recovery	=	93.70%	
35) Dibromofluoromethane	7.56	113	100163	49.21	ug/l	0.00
Spiked Amount			Recovery	=	98.42%	
50) Toluene-d8	10.08	98	416195	50.06	ug/l	0.00
Spiked Amount			Recovery	=	100.12%	
62) 4-Bromofluorobenzene	12.40	95	157492	51.69	ug/l	0.00
Spiked Amount			Recovery	=	103.38%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.72	59	189	0.462	ug/l #	72
16) Acetone	3.78	43	40606	41.219	ug/l	100
18) Methyl Acetate	4.30	43	2186	0.788	ug/l #	60
20) Methylene Chloride	4.52	84	14002	5.778	ug/l	97
25) 2-Butanone	6.80	43	15358	9.634	ug/l	96
29) Tetrahydrofuran	7.19	42	348	0.309	ug/l #	35
31) Cyclohexane	7.64	56	4185	0.934	ug/l #	1
36) 1,1-Dichloropropene	7.64	75	15553	4.772	ug/l #	49
37) Ethyl Acetate	6.80	43	15358	4.361	ug/l #	70
38) Carbon Tetrachloride	7.64	117	19005	6.506	ug/l #	16
40) Benzene	8.02	78	4028	0.409	ug/l	95
43) Isopropyl Acetate	8.14	43	56428	9.716	ug/l #	91
45) 1,2-Dichloropropane	9.25	63	927	0.357	ug/l #	43
48) Methyl methacrylate	9.16	41	4629	1.790	ug/l #	32
51) 4-Methyl-2-Pentanone	9.97	43	117025	35.409	ug/l	97
52) Toluene	10.14	92	4127	0.695	ug/l	97
54) cis-1,3-Dichloropropene	9.89	75	20749	5.048	ug/l #	61
56) Ethyl methacrylate	10.34	69	6053	1.698	ug/l #	68
57) 1,3-Dichloropropane	10.75	76	3099	0.781	ug/l #	42
59) 2-Hexanone	10.75	43	46957	19.724	ug/l #	54
71) Bromoform	12.39	173	337	1.852	ug/l #	1
74) N-amyl acetate	12.02	43	6302	1.217	ug/l #	48
75) 1,1,2,2-Tetrachloroethane	12.56	83	1151	0.340	ug/l #	28
76) 1,2,3-Trichloropropane	12.40	75	78837	24.470	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	78837	62.847	ug/l #	13
95) Naphthalene	15.13	128	4498	0.535	ug/l #	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060858.D
Acq On : 6 Apr 2020 17:50
Operator : JC/MD
Sample : L2144-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH

Quant Time: Apr 07 07:06:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :


CHASSIS-WASH

Tot Ion:168 Resp: 202590

Ion Ratio Lower Upper




168 100

99 56.1 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#11

Tert butyl alcohol

Concen: 0.462 ug/l

RT: 4.72 min Scan# 915

Delta R.T. -0.01 min

Lab File: VN060858.D


Acq: 6 Apr 2020 17:50

Tgt Ion: 59 Resp: 189

Ion Ratio Lower Upper



59 100


57 0.0 8.2 12.4#



Abundance Ion 59.00 (58.70 to 59.70): VN0

Ion 57.00 (56.70 to 57.70): VN0





#16

Acetone

Concen: 41.219 ug/l

RT: 3.78 min Scan# 625

Delta R.T. 0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :





CHASSIS-WASH

Tot Ion: 43 Resp: 40606

Ion Ratio Lower Upper

43 100

58 31.5 25.1 37.7



#18

Methyl Acetate

Concen: 0.788 ug/l

RT: 4.30 min Scan# 785

Delta R.T. 0.01 min

Lab File: VN060858.D




Acq: 6 Apr 2020 17:50


Tgt Ion: 43 Resp: 2186

Ion Ratio Lower Upper

43 100

74 2.9 17.5 26.3#









#20
Methvlene Chloride
Concen: 5.778 ug/l
RT: 4.52 min Scan# 855
Delta R.T. 0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH

Tot Ion:	84	Resp:	14002
Ion	Ratio	Lower	Upper
84	100		
49	134.4	103.3	154.9
51	39.7	31.6	47.4
86	62.5	52.0	78.0




Abundance Ion 83.90 (83.60 to 84.60): VN0
Ion 48.90 (48.60 to 49.60): VN0
Ion 50.90 (50.60 to 51.60): VN0
Ion 85.90 (85.60 to 86.60): VN0






#25
2-Butanone
Concen: 9.634 ug/l
RT: 6.80 min Scan# 1564
Delta R.T. 0.01 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion:	43	Resp:	15358
Ion	Ratio	Lower	Upper
43	100		
72	23.6	20.4	30.6



Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 72.00 (71.70 to 72.70): VN0









#29
Tetrahydrofuran
Concen: 0.309 ug/l
RT: 7.19 min Scan# 1684
Delta R.T. 0.01 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50




Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH


Tot Ion: 42 Resp: 348
Ion Ratio Lower Upper
42 100
72 0.0 33.2 49.8#
71 0.0 31.1 46.7#



#31
Cyclohexane
Concen: 0.934 ug/l
RT: 7.64 min Scan# 1826
Delta R.T. 0.01 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion: 56 Resp: 4185
Ion Ratio Lower Upper
56 100
69 168.8 26.6 40.0#
84 148.6 64.3 96.5#









#33
 1,2-Dichloroethane-d4
 Concen: 46.854 ug/l
 RT: 8.00 min Scan# 1937
 Delta R.T. -0.00 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50




Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH


Tot Ion: 65 Resp: 130305
 Ion Ratio Lower Upper
 65 100
 67 53.6 0.0 107.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. 0.00 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50

Tgt Ion: 114 Resp: 336587
 Ion Ratio Lower Upper
 114 100
 63 22.0 0.0 43.8
 88 16.4 0.0 33.0









#35
 Dibromofluoromethane
 Concen: 49.214 ug/l
 RT: 7.56 min Scan# 1801
 Delta R.T. 0.00 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50




Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH


Tot Ion	Ratio	Lower	Upper
113	100		
111	100.9	82.9	124.3
192	19.8	15.8	23.8



#36
 1,1-Dichloropropene
 Concen: 4.772 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.13 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50

Tgt Ion	Ratio	Lower	Upper
75	100		
110	7.4	16.9	50.6#
77	0.0	24.6	36.8#









#37
Ethyl Acetate
Concen: 4.361 ug/l
RT: 6.80 min Scan# 1564
Delta R.T. -0.09 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50




Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH


Tot Ion: 43 Resp: 15358
Ion Ratio Lower Upper
43 100
61 0.0 10.6 15.8#
70 0.0 7.9 11.9#



#38
Carbon Tetrachloride
Concen: 6.506 ug/l
RT: 7.64 min Scan# 1825
Delta R.T. -0.11 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion: 117 Resp: 19005
Ion Ratio Lower Upper
117 100
119 4.1 76.2 114.2#
121 0.0 23.7 35.5#





#40

Benzene

Concen: 0.409 ug/l

RT: 8.02 min Scan# 1942

Delta R.T. -0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :





CHASSIS-WASH

Tot Ion: 78 Resp: 4028

Ion Ratio Lower Upper

78 100

77 25.7 18.6 28.0



#43

Isopropyl Acetate

Concen: 9.716 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50




Tgt Ion: 43 Resp: 56428


Ion Ratio Lower Upper

43 100

61 19.4 20.7 31.1#

87 11.6 9.8 14.8









#45
 1,2-Dichloropropane
 Concen: 0.357 ug/l
 RT: 9.25 min Scan# 2324
 Delta R.T. 0.14 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50




Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH


Tot Ion: 63 Resp: 927
 Ion Ratio Lower Upper
 63 100
 65 0.0 25.4 38.0#



#48
 Methyl methacrylate
 Concen: 1.790 ug/l
 RT: 9.16 min Scan# 2296
 Delta R.T. -0.03 min
 Lab File: VN060858.D
 Acq: 6 Apr 2020 17:50

Tgt Ion: 41 Resp: 4629
 Ion Ratio Lower Upper
 41 100
 69 0.0 65.4 98.0#
 39 30.9 44.0 66.0#









#50
Toluene-d8
Concen: 50.065 ug/l
RT: 10.08 min Scan# 2583
Delta R.T. 0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50




Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH


Tot Ion: 98 Resp: 416195
Ion Ratio Lower Upper
98 100
100 65.2 51.8 77.8



#51
4-Methyl-2-Pentanone
Concen: 35.409 ug/l
RT: 9.97 min Scan# 2548
Delta R.T. -0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion: 43 Resp: 117025
Ion Ratio Lower Upper
43 100
58 36.7 30.8 46.2





#52

Toluene

Concen: 0.695 ug/l

RT: 10.14 min Scan# 2602

Delta R.T. -0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :


CHASSIS-WASH

Tot Ion: 92 Resp: 4127

Ion Ratio Lower Upper




92 100

91 176.5 137.6 206.4



Abundance Ion 92.00 (91.70 to 92.70): VN0

Ion 91.00 (90.70 to 91.70): VN0



#54

cis-1,3-Dichloropropene

Concen: 5.048 ug/l

RT: 9.89 min Scan# 2524

Delta R.T. 0.06 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50


Tgt Ion: 75 Resp: 20749

Ion Ratio Lower Upper

75 100

77 0.0 25.4 38.0#



39 30.9 40.2 60.4#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 76.90 (76.60 to 77.60): VN0

Ion 39.00 (38.70 to 39.70): VN0





#56

Ethyl methacrylate

Concen: 1.698 ug/l

RT: 10.34 min Scan# 2665

Delta R.T. -0.07 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :

CHASSIS-WASH


Tot Ion: 69 Resp: 6053

Ion Ratio Lower Upper

69 100

41 95.9 55.4 83.2#




39 55.8 29.4 44.0#



Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 0.781 ug/l

RT: 10.75 min Scan# 2793

Delta R.T. 0.06 min

Lab File: VN060858.D


Acq: 6 Apr 2020 17:50

Tgt Ion: 76 Resp: 3099

Ion Ratio Lower Upper



76 100


78 0.0 25.8 38.6#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0









#59
2-Hexanone
Concen: 19.724 ug/l
RT: 10.75 min Scan# 2793
Delta R.T. 0.01 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50




Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH


Tot Ion: 43 Resp: 46957
Ion Ratio Lower Upper
43 100
58 20.5 26.7 80.1#



#62
4-Bromofluorobenzene
Concen: 51.694 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion: 95 Resp: 157492
Ion Ratio Lower Upper
95 100
174 77.8 0.0 157.6
176 76.0 0.0 150.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH

Tot Ion:117 Resp: 317462
Ion Ratio Lower Upper
117 100
82 57.9 46.7 70.1
119 32.3 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 1.852 ug/l
RT: 12.39 min Scan# 3303
Delta R.T. 0.27 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion:173 Resp: 337
Ion Ratio Lower Upper
173 100
175 2370.6 24.1 72.4#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :

CHASSIS-WASH


Tot Ion:152 Resp: 149308

Ion Ratio Lower Upper

152 100

115 59.4 30.2 90.6




150 155.7 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 1.217 ug/l

RT: 12.02 min Scan# 3188

Delta R.T. -0.04 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Tgt Ion: 43 Resp: 6302


Ion Ratio Lower Upper

43 100

70 0.0 31.1 46.7#

55 6.4 18.6 27.8#

61 0.0 18.9 28.3#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0





#75

1,1,2,2-Tetrachloroethane

Concen: 0.340 ug/l

RT: 12.56 min Scan# 3356

Delta R.T. 0.06 min

Lab File: VN060858.D

Acq: 6 Apr 2020 17:50

Instrument :

MSVOA_N

ClientSampleId :

CHASSIS-WASH


Tot Ion: 83 Resp: 1151

Ion Ratio Lower Upper

83 100

131 0.0 4.9 14.7#

85 2.6 32.5 97.4#



Abundance Ion 82.90 (82.60 to 83.60): VN0

Ion 130.80 (130.50 to 131.50): VN0

Ion 84.90 (84.60 to 85.60): VN0

12.56

600

400



200

0

12.55

12.60

Time-->



#76

1,2,3-Trichloropropane

Concen: 24.470 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.15 min

Lab File: VN060858.D


Acq: 6 Apr 2020 17:50

Tgt Ion: 75 Resp: 78837

Ion Ratio Lower Upper

75 100

77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0

12.40

50000

40000

30000

20000

10000


0


12.35

12.40

12.45

Time-->









#81
trans-1,4-Dichloro-2-butene
Concen: 62.847 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.10 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH

Tot Ion: 75 Resp: 78837
Ion Ratio Lower Upper
75 100
53 0.0 73.0 109.4#
89 0.0 34.8 52.2#



#95
Naphthalene
Concen: 0.535 ug/l
RT: 15.13 min Scan# 4154
Delta R.T. 0.00 min
Lab File: VN060858.D
Acq: 6 Apr 2020 17:50

Tgt Ion: 128 Resp: 4498
Ion Ratio Lower Upper
128 100
127 11.6 10.2 15.2
129 7.6 8.8 13.2#

