

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060848.D
 Acq On : 6 Apr 2020 14:04
 Operator : JC/MD
 Sample : L2129-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Apr 07 08:05:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	209890	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	347758	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321784	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142765	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	132969	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
35) Dibromofluoromethane	7.57	113	102556	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.08	98	427585	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
62) 4-Bromofluorobenzene	12.40	95	158726	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
Target Compounds						
16) Acetone	3.79	43	6839	6.701	ug/l	98
20) Methylene Chloride	4.53	84	9073	3.614	ug/l	97
43) Isopropyl Acetate	8.14	43	55044	9.173	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060848.D
Acq On : 6 Apr 2020 14:04
Operator : JC/MD
Sample : L2129-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Apr 07 08:05:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

