

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060850.D
 Acq On : 6 Apr 2020 14:50
 Operator : JC/MD
 Sample : L2129-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Apr 07 08:08:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	217208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	358968	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	334517	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	138517	46.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.90%	
35) Dibromofluoromethane	7.56	113	106173	48.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.82%	
50) Toluene-d8	10.08	98	440153	49.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.30%	
62) 4-Bromofluorobenzene	12.40	95	163287	50.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.50%	
Target Compounds						
16) Acetone	3.78	43	8291	7.850	ug/l	95
20) Methylene Chloride	4.52	84	7745	2.981	ug/l	94
43) Isopropyl Acetate	8.14	43	54727	8.835	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060850.D
Acq On : 6 Apr 2020 14:50
Operator : JC/MD
Sample : L2129-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Apr 07 08:08:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

