

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060852.D
 Acq On : 6 Apr 2020 15:35
 Operator : JC/MD
 Sample : L2129-34
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Quant Time: Apr 07 08:12:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	213349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	354088	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	331100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134899	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
35) Dibromofluoromethane	7.56	113	105512	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	10.08	98	437398	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
62) 4-Bromofluorobenzene	12.40	95	161174	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
Target Compounds						
16) Acetone	3.79	43	7355	7.090	ug/l	100
20) Methylene Chloride	4.53	84	6412	2.512	ug/l	91
43) Isopropyl Acetate	8.14	43	53330	8.729	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060852.D
Acq On : 6 Apr 2020 15:35
Operator : JC/MD
Sample : L2129-34
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Apr 07 08:12:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

