

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060855.D
 Acq On : 6 Apr 2020 16:43
 Operator : JC/MD
 Sample : L2132-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T-1

Quant Time: Apr 07 08:20:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	329604	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	308182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	139073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	127539	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
35) Dibromofluoromethane	7.56	113	97659	49.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.00%	
50) Toluene-d8	10.08	98	406791	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	12.40	95	149424	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
Target Compounds						
16) Acetone	3.78	43	40970	42.416	ug/l	98
20) Methylene Chloride	4.53	84	7299	3.072	ug/l	93
25) 2-Butanone	6.80	43	11894	7.610	ug/l	94
43) Isopropyl Acetate	8.14	43	11648	2.048	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060855.D
Acq On : 6 Apr 2020 16:43
Operator : JC/MD
Sample : L2132-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T-1

Quant Time: Apr 07 08:20:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

