

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040620\
 Data File : VN060865.D
 Acq On : 6 Apr 2020 20:29
 Operator : JC/MD
 Sample : L2162-13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-50-COMP

Quant Time: Apr 07 08:58:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	211677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	350862	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	326786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	150971	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134398	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
35) Dibromofluoromethane	7.56	113	103222	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	10.08	98	429879	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	12.40	95	162863	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
Target Compounds						
16) Acetone	3.78	43	11479	11.152	ug/l	92
20) Methylene Chloride	4.52	84	7488	2.957	ug/l	97
43) Isopropyl Acetate	8.14	43	52851	8.730	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040620\
Data File : VN060865.D
Acq On : 6 Apr 2020 20:29
Operator : JC/MD
Sample : L2162-13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-50-COMP

Quant Time: Apr 07 08:58:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

