

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
 Data File : VN066498.D
 Acq On : 6 Apr 2021 17:51
 Operator : JC/MD
 Sample : PB135551TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135551TB

Quant Time: Apr 07 04:47:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	316009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	470781	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	463635	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	200094	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	193273	46.44	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.88%
35) Dibromofluoromethane	7.507	113	154970	48.30	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.60%
50) Toluene-d8	10.028	98	604119	49.55	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.10%
62) 4-Bromofluorobenzene	12.348	95	187717	47.48	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.96%
Target Compounds						
						Qvalue
16) Acetone	3.741	43	24092	13.81	ug/l	98
20) Methylene Chloride	4.457	84	11550	3.08	ug/l	94
43) Isopropyl Acetate	8.089	43	31233	3.56	ug/l #	86
95) Naphthalene	15.274	128	19997	3.61	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
Data File : VN066498.D
Acq On : 6 Apr 2021 17:51
Operator : JC/MD
Sample : PB135551TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135551TB

Quant Time: Apr 07 04:47:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

