

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
 Data File : VN066499.D
 Acq On : 6 Apr 2021 18:14
 Operator : JC/MD
 Sample : PB135551ZHE#01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135551ZHE#01

Quant Time: Apr 07 04:48:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	343112	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	505720	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	480442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	189614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	212218	46.97	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.94%
35) Dibromofluoromethane	7.507	113	167390	48.56	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.12%
50) Toluene-d8	10.028	98	642521	49.06	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.12%
62) 4-Bromofluorobenzene	12.345	95	191386	45.07	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.14%
Target Compounds						
					Qvalue	
16) Acetone	3.733	43	23383	12.35	ug/l	94
20) Methylene Chloride	4.452	84	12190	3.00	ug/l	97
43) Isopropyl Acetate	8.089	43	31462	3.34	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
Data File : VN066499.D
Acq On : 6 Apr 2021 18:14
Operator : JC/MD
Sample : PB135551ZHE#01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135551ZHE#01

Quant Time: Apr 07 04:48:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

