

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
 Data File : VN066502.D
 Acq On : 6 Apr 2021 19:25
 Operator : JC/MD
 Sample : M1934-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Apr 07 04:49:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	319104	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	475780	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	464683	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	205878	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	198676	47.28	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.56%
35) Dibromofluoromethane	7.504	113	156019	48.11	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.22%
50) Toluene-d8	10.028	98	610824	49.58	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.16%
62) 4-Bromofluorobenzene	12.348	95	193676	48.47	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.94%
Target Compounds						
					Qvalue	
16) Acetone	3.739	43	24514	13.92	ug/l	92
20) Methylene Chloride	4.460	84	12838	3.40	ug/l	90
43) Isopropyl Acetate	8.089	43	29256	3.30	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040621\
Data File : VN066502.D
Acq On : 6 Apr 2021 19:25
Operator : JC/MD
Sample : M1934-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Apr 07 04:49:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

