

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071882.D
 Acq On : 06 Apr 2022 16:19
 Operator : JC\MD
 Sample : PB143897ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143897ZHE#02

Quant Time: Apr 07 04:28:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	605273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1056015	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1096703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	405280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	337730	48.876	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.760%
35) Dibromofluoromethane	8.016	113	297524	47.879	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.760%
50) Toluene-d8	10.439	98	1343649	52.858	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.720%
62) 4-Bromofluorobenzene	12.727	95	454056	55.229	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.460%
Target Compounds						
					Qvalue	
16) Acetone	4.298	43	19002	6.435	ug/l	94
20) Methylene Chloride	5.098	84	39144	6.289	ug/l	94
43) Isopropyl Acetate	8.569	43	128889	8.159	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
Data File : VN071882.D
Acq On : 06 Apr 2022 16:19
Operator : JC\MD
Sample : PB143897ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143897ZHE#02

Quant Time: Apr 07 04:28:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

