

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071883.D
 Acq On : 06 Apr 2022 16:43
 Operator : JC\MD
 Sample : PB143897ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143897ZHE#03

Quant Time: Apr 07 04:28:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	624566	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1091400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1125229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	406933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	344414	48.304	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.600%
35) Dibromofluoromethane	8.022	113	304486	47.411	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.820%
50) Toluene-d8	10.439	98	1389299	52.881	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.760%
62) 4-Bromofluorobenzene	12.727	95	465160	54.746	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.500%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	18147	5.955	ug/l	100
20) Methylene Chloride	5.098	84	38850	6.049	ug/l	95
43) Isopropyl Acetate	8.575	43	129106	7.907	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
Data File : VN071883.D
Acq On : 06 Apr 2022 16:43
Operator : JC\MD
Sample : PB143897ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143897ZHE#03

Quant Time: Apr 07 04:28:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

