

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071887.D
 Acq On : 06 Apr 2022 18:20
 Operator : JC\MD
 Sample : N2254-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-4-GR-5

Quant Time: Apr 07 04:29:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	607803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1054886	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1108000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	400559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	332807	47.963	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.920%
35) Dibromofluoromethane	8.022	113	296642	47.788	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.580%
50) Toluene-d8	10.439	98	1346354	53.021	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.727	95	454935	55.395	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.800%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	20580	6.940	ug/l	100
20) Methylene Chloride	5.098	84	34498	5.519	ug/l	94
43) Isopropyl Acetate	8.569	43	152083	9.637	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071887.D
 Acq On : 06 Apr 2022 18:20
 Operator : JC\MD
 Sample : N2254-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A-1A-4-GR-5

Quant Time: Apr 07 04:29:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

