

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071890.D
 Acq On : 06 Apr 2022 19:32
 Operator : JC\MD
 Sample : N2272-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Apr 07 04:30:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	550561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	947629	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	983341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	342291	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	306285	48.730	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.460%
35) Dibromofluoromethane	8.022	113	268269	48.109	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.220%
50) Toluene-d8	10.439	98	1215746	53.296	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.600%
62) 4-Bromofluorobenzene	12.727	95	395331	53.586	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.180%
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	23942	8.913	ug/l	94
18) Methyl Acetate	4.869	43	23021	2.008	ug/l	99
20) Methylene Chloride	5.098	84	27306	4.823	ug/l #	90
43) Isopropyl Acetate	8.575	43	242246	17.088	ug/l #	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
Data File : VN071890.D
Acq On : 06 Apr 2022 19:32
Operator : JC\MD
Sample : N2272-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Apr 07 04:30:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

