

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
 Data File : VN071891.D
 Acq On : 06 Apr 2022 19:56
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Apr 07 04:30:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	580851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1024931	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1058256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	397831	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	323972	48.857	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.720%
35) Dibromofluoromethane	8.022	113	286193	47.453	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.900%
50) Toluene-d8	10.439	98	1267618	51.379	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.760%
62) 4-Bromofluorobenzene	12.727	95	441236	55.298	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.600%
Target Compounds						
					Qvalue	
18) Methyl Acetate	4.863	43	23480	1.889	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040622\
Data File : VN071891.D
Acq On : 06 Apr 2022 19:56
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Apr 07 04:30:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

