

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060875.D
 Acq On : 7 Apr 2020 14:33
 Operator : JC/MD
 Sample : L2178-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Apr 08 07:57:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	186542	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	308646	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	284590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123986	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119360	46.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.22%	
35) Dibromofluoromethane	7.56	113	91966	49.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.56%	
50) Toluene-d8	10.08	98	381163	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	12.40	95	136416	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
Target Compounds						
16) Acetone	3.79	43	3830	4.222	ug/l	97
20) Methylene Chloride	4.52	84	2483	1.113	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060875.D
Acq On : 7 Apr 2020 14:33
Operator : JC/MD
Sample : L2178-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Apr 08 07:57:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

