

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060877.D
 Acq On : 7 Apr 2020 15:18
 Operator : JC/MD
 Sample : L2167-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 08 08:12:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	197777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	333595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	310791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136464	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	129105	47.55	ug/l	0.00
Spiked Amount			Recovery	=	95.10%	
35) Dibromofluoromethane	7.56	113	100049	49.60	ug/l	0.00
Spiked Amount			Recovery	=	99.20%	
50) Toluene-d8	10.08	98	409612	49.71	ug/l	0.00
Spiked Amount			Recovery	=	99.42%	
62) 4-Bromofluorobenzene	12.40	95	150594	49.87	ug/l	0.00
Spiked Amount			Recovery	=	99.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060877.D
Acq On : 7 Apr 2020 15:18
Operator : JC/MD
Sample : L2167-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 08 08:12:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

