

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060880.D
 Acq On : 7 Apr 2020 16:26
 Operator : JC/MD
 Sample : L2166-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-47-COMP

Quant Time: Apr 08 08:20:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	200782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	337285	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	316260	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142880	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131277	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	7.56	113	99068	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	10.08	98	415128	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
62) 4-Bromofluorobenzene	12.40	95	154981	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
Target Compounds						
16) Acetone	3.78	43	10690	10.949	ug/l	98
20) Methylene Chloride	4.53	84	3531	1.470	ug/l	98
43) Isopropyl Acetate	8.14	43	54659	9.392	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060880.D
Acq On : 7 Apr 2020 16:26
Operator : JC/MD
Sample : L2166-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-47-COMP

Quant Time: Apr 08 08:20:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

