

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060886.D
 Acq On : 7 Apr 2020 18:42
 Operator : JC/MD
 Sample : L2178-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Quant Time: Apr 08 08:37:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	331750	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	306547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	135746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	128253	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	7.57	113	99050	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	10.08	98	407662	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.40	95	148997	49.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.24%	
Target Compounds						
16) Acetone	3.79	43	3599	3.721	ug/l	92
20) Methylene Chloride	4.52	84	1689	0.710	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060886.D
Acq On : 7 Apr 2020 18:42
Operator : JC/MD
Sample : L2178-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Quant Time: Apr 08 08:37:03 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

