

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060888.D
 Acq On : 7 Apr 2020 19:27
 Operator : JC/MD
 Sample : L2143-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FORMER-OLDEN-AVE-SUB

Quant Time: Apr 08 08:42:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195955	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	327266	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	307279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	140010	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	126371	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
35) Dibromofluoromethane	7.56	113	96825	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
50) Toluene-d8	10.08	98	401939	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
62) 4-Bromofluorobenzene	12.40	95	149682	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
Target Compounds						
16) Acetone	3.78	43	8945	9.388	ug/l	98
17) Carbon Disulfide	4.03	76	6464	1.002	ug/l	99
20) Methylene Chloride	4.52	84	14935	6.372	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060888.D
Acq On : 7 Apr 2020 19:27
Operator : JC/MD
Sample : L2143-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FORMER-OLDEN-AVE-SUB

Quant Time: Apr 08 08:42:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

