

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN040720\
 Data File : VN060877.D
 Acq On : 7 Apr 2020 15:18
 Operator : JC/MD
 Sample : L2167-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 08 08:12:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	197777	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	333595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	310791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136464	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	129105	47.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.10%	
35) Dibromofluoromethane	7.56	113	100049	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
50) Toluene-d8	10.08	98	409612	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
62) 4-Bromofluorobenzene	12.40	95	150594	49.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.74%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN040720\
Data File : VN060877.D
Acq On : 7 Apr 2020 15:18
Operator : JC/MD
Sample : L2167-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Quant Time: Apr 08 08:12:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.00 min

Lab File: VN060877.D

Acq: 7 Apr 2020 15:18

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-SOIL-REF-5

Tot Ion:168 Resp: 197777

Ion Ratio Lower Upper

168 100

99 56.3 44.0 66.0



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.64

80000

60000



40000

20000

0

7.60 7.70

Time-->



#33

1,2-Dichloroethane-d4

Concen: 47.552 ug/l

RT: 8.00 min Scan# 1938

Delta R.T. 0.00 min

Lab File: VN060877.D


Acq: 7 Apr 2020 15:18

Tgt Ion: 65 Resp: 129105

Ion Ratio Lower Upper

65 100

67 53.8 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0

8.00

60000

50000

40000

30000


20000


10000

0

7.90 8.00 8.10

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.56 min Scan# 2112
 Delta R.T. -0.00 min
 Lab File: VN060877.D
 Acq: 7 Apr 2020 15:18

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-5

Tot Ion	Ratio	Lower	Upper
114	100		
63	22.2	0.0	43.8
88	16.8	0.0	33.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 49.599 ug/l
 RT: 7.56 min Scan# 1801
 Delta R.T. 0.00 min
 Lab File: VN060877.D
 Acq: 7 Apr 2020 15:18

Tgt Ion	Ratio	Lower	Upper
113	100		
111	102.7	82.9	124.3
192	20.0	15.8	23.8



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.715 ug/l

RT: 10.08 min Scan# 2582

Delta R.T. -0.00 min

Lab File: VN060877.D

Acq: 7 Apr 2020 15:18

Instrument :

MSVOA_N

ClientSampleId :





STORAGE-BLANK-SOIL-REF-5

Tot Ion: 98 Resp: 409612

Ion Ratio Lower Upper

98 100

100 65.2 51.8 77.8



#62

4-Bromofluorobenzene

Concen: 49.873 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN060877.D

Acq: 7 Apr 2020 15:18




Tgt Ion: 95 Resp: 150594

Ion Ratio Lower Upper

95 100

174 78.7 0.0 157.6

176 75.5 0.0 150.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060877.D
Acq: 7 Apr 2020 15:18

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SOIL-REF-5

Tot Ion:117 Resp: 310791
Ion Ratio Lower Upper
117 100
82 58.2 46.7 70.1
119 32.6 25.5 38.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3597
Delta R.T. 0.00 min
Lab File: VN060877.D
Acq: 7 Apr 2020 15:18

Tgt Ion:152 Resp: 136464
Ion Ratio Lower Upper
152 100
115 60.1 30.2 90.6
150 157.5 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

