

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040721\  
 Data File : VN066508.D  
 Acq On : 7 Apr 2021 16:05  
 Operator : JC/MD  
 Sample : VN0407WBL01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 VN0407WBL01

Quant Time: Apr 08 01:54:20 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N040621W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 06 15:38:34 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units     | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-----------|----------|
| -----                       |        |       |          |          |           |          |
| Internal Standards          |        |       |          |          |           |          |
| 1) Pentafluorobenzene       | 7.582  | 168   | 367099   | 50.00    | ug/l      | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.512  | 114   | 555249   | 50.00    | ug/l      | 0.00     |
| 63) Chlorobenzene-d5        | 11.347 | 117   | 515895   | 50.00    | ug/l      | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.289 | 152   | 226993   | 50.00    | ug/l      | 0.00     |
| System Monitoring Compounds |        |       |          |          |           |          |
| 33) 1,2-Dichloroethane-d4   | 7.947  | 65    | 263220   | 54.45    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | = 108.90% |          |
| 35) Dibromofluoromethane    | 7.507  | 113   | 207069   | 54.72    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 109.44% |          |
| 50) Toluene-d8              | 10.025 | 98    | 779721   | 54.23    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | = 108.46% |          |
| 62) 4-Bromofluorobenzene    | 12.345 | 95    | 237668   | 50.97    | ug/l      | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 101.94% |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040721\  
Data File : VN066508.D  
Acq On : 7 Apr 2021 16:05  
Operator : JC/MD  
Sample : VN0407WBL01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
VN0407WBL01

Quant Time: Apr 08 01:54:20 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N040621W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 06 15:38:34 2021  
Response via : Initial Calibration

