

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066511.D
 Acq On : 7 Apr 2021 17:25
 Operator : JC/MD
 Sample : PB135583TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583TB

Quant Time: Apr 08 01:55:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	331010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.513	114	502007	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	489881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	204077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	208394	47.81	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.62%
35) Dibromofluoromethane	7.507	113	162922	47.62	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.24%
50) Toluene-d8	10.028	98	638724	49.13	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.26%
62) 4-Bromofluorobenzene	12.345	95	199721	47.38	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.76%
Target Compounds						
16) Acetone	3.741	43	23029	12.60	ug/l	95
20) Methylene Chloride	4.457	84	13075	3.33	ug/l	89
43) Isopropyl Acetate	8.089	43	32197	3.44	ug/l	97
95) Naphthalene	15.271	128	27534	4.18	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066511.D
Acq On : 7 Apr 2021 17:25
Operator : JC/MD
Sample : PB135583TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583TB

Quant Time: Apr 08 01:55:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

