

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040721\  
 Data File : VN066512.D  
 Acq On : 7 Apr 2021 17:48  
 Operator : JC/MD  
 Sample : PB135583ZH#01  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 PB135583ZH#01

Quant Time: Apr 08 01:56:11 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N040621W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 06 15:38:34 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 7.584  | 168   | 324396   | 50.00    | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.515  | 114   | 490457   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.350 | 117   | 485342   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.289 | 152   | 214949   | 50.00    | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 7.947  | 65    | 205518   | 48.11    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 96.22%   |
| 35) Dibromofluoromethane    | 7.509  | 113   | 160101   | 47.89    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 95.78%   |
| 50) Toluene-d8              | 10.028 | 98    | 626991   | 49.37    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 98.74%   |
| 62) 4-Bromofluorobenzene    | 12.348 | 95    | 198608   | 48.22    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 96.44%   |
| Target Compounds            |        |       |          |          |       |          |
| 16) Acetone                 | 3.744  | 43    | 22616    | 12.63    | ug/l  | 97       |
| 20) Methylene Chloride      | 4.463  | 84    | 16874    | 4.39     | ug/l  | 85       |
| 43) Isopropyl Acetate       | 8.089  | 43    | 30566    | 3.35     | ug/l  | 100      |
| -----                       |        |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN040721\  
Data File : VN066512.D  
Acq On : 7 Apr 2021 17:48  
Operator : JC/MD  
Sample : PB135583ZH#01  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
PB135583ZH#01

Quant Time: Apr 08 01:56:11 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N040621W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 06 15:38:34 2021  
Response via : Initial Calibration

