

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066514.D
 Acq On : 7 Apr 2021 18:35
 Operator : JC/MD
 Sample : PB135583ZH#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#03

Quant Time: Apr 08 01:56:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	322570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	490843	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	489435	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	221517	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	208100	48.99	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.98%
35) Dibromofluoromethane	7.507	113	164060	49.04	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.08%
50) Toluene-d8	10.028	98	636061	50.04	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.08%
62) 4-Bromofluorobenzene	12.348	95	195788	47.50	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.00%
Target Compounds						
16) Acetone	3.733	43	22758	12.78	ug/l	99
20) Methylene Chloride	4.452	84	13044	3.41	ug/l	98
43) Isopropyl Acetate	8.086	43	30844	3.37	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066514.D
Acq On : 7 Apr 2021 18:35
Operator : JC/MD
Sample : PB135583ZH#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#03

Quant Time: Apr 08 01:56:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

