

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066516.D
 Acq On : 7 Apr 2021 19:23
 Operator : JC/MD
 Sample : PB135583ZH#05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#05

Quant Time: Apr 08 01:57:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	357366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	541966	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	531310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	221438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	228270	48.50	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.00%		
35) Dibromofluoromethane	7.507	113	178402	48.30	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.60%		
50) Toluene-d8	10.028	98	698492	49.77	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.54%		
62) 4-Bromofluorobenzene	12.345	95	211230	46.41	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.82%		
Target Compounds						Qvalue
16) Acetone	3.738	43	21194	10.74	ug/l	94
20) Methylene Chloride	4.455	84	11894	2.81	ug/l	93
43) Isopropyl Acetate	8.091	43	30794	3.05	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066516.D
Acq On : 7 Apr 2021 19:23
Operator : JC/MD
Sample : PB135583ZH#05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#05

Quant Time: Apr 08 01:57:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

