

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066517.D
 Acq On : 7 Apr 2021 19:46
 Operator : JC/MD
 Sample : PB135583ZH#06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#06

Quant Time: Apr 08 01:57:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	320565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	485034	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	495522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	215638	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	209297	49.58	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.16%
35) Dibromofluoromethane	7.507	113	164852	49.87	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.74%
50) Toluene-d8	10.028	98	628787	50.06	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.12%
62) 4-Bromofluorobenzene	12.345	95	203898	50.06	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.12%
Target Compounds						
16) Acetone	3.739	43	21741	12.29	ug/l	95
18) Methyl Acetate	4.243	43	4731	1.07	ug/l #	70
20) Methylene Chloride	4.457	84	16176	4.26	ug/l	91
43) Isopropyl Acetate	8.089	43	30543	3.38	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066517.D
Acq On : 7 Apr 2021 19:46
Operator : JC/MD
Sample : PB135583ZH#06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#06

Quant Time: Apr 08 01:57:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

