

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066519.D
 Acq On : 7 Apr 2021 20:33
 Operator : JC/MD
 Sample : PB135583ZH#08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#08

Quant Time: Apr 08 01:58:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	316530	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	474218	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	466821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	204748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	203845	48.90	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	97.80%	
35) Dibromofluoromethane	7.504	113	161308	49.91	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.82%	
50) Toluene-d8	10.028	98	619552	50.45	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.90%	
62) 4-Bromofluorobenzene	12.345	95	189141	47.49	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	94.98%	
Target Compounds						
16) Acetone	3.741	43	23043	13.19	ug/l	97
18) Methyl Acetate	4.237	43	4620m	1.06	ug/l	
20) Methylene Chloride	4.457	84	16483	4.39	ug/l	92
43) Isopropyl Acetate	8.091	43	30914	3.50	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066519.D
Acq On : 7 Apr 2021 20:33
Operator : JC/MD
Sample : PB135583ZH#08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#08

Quant Time: Apr 08 01:58:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

