

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066520.D
 Acq On : 7 Apr 2021 20:57
 Operator : JC/MD
 Sample : PB135583ZH#09
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#09

Quant Time: Apr 08 01:59:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	300730	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	458768	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	459667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	193587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	194753	49.17	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.34%		
35) Dibromofluoromethane	7.507	113	153273	49.02	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.04%		
50) Toluene-d8	10.028	98	591772	49.81	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.62%		
62) 4-Bromofluorobenzene	12.345	95	185868	48.24	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.48%		
Target Compounds						
						Qvalue
16) Acetone	3.736	43	23238	14.00	ug/l	99
18) Methyl Acetate	4.240	43	4212	1.01	ug/l #	88
20) Methylene Chloride	4.457	84	16619	4.66	ug/l	93
43) Isopropyl Acetate	8.089	43	30386	3.56	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066520.D
Acq On : 7 Apr 2021 20:57
Operator : JC/MD
Sample : PB135583ZH#09
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#09

Quant Time: Apr 08 01:59:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

