

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066523.D
 Acq On : 7 Apr 2021 22:07
 Operator : JC/MD
 Sample : PB135583ZH#12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#12

Manual Integrations
 APPROVED

MMDadoda
 4/9/2021 12:49:27 PM

Quant Time: Apr 08 02:00:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	347282	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	524244	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	510511	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	219440	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.949	65	219603	48.02	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.04%		
35) Dibromofluoromethane	7.509	113	175462	49.11	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.22%		
50) Toluene-d8	10.028	98	671457	49.46	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.92%		
62) 4-Bromofluorobenzene	12.348	95	209836	47.66	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.32%		
Target Compounds						Qvalue
16) Acetone	3.738	43	23033	12.02	ug/l	98
18) Methyl Acetate	4.237	43	5968m	1.24	ug/l	
20) Methylene Chloride	4.454	84	17574	4.27	ug/l	87
43) Isopropyl Acetate	8.086	43	30359	3.11	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066523.D
Acq On : 7 Apr 2021 22:07
Operator : JC/MD
Sample : PB135583ZH#12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#12

Manual Integrations
APPROVED

MMDadoda
4/9/2021 12:49:27 PM

Quant Time: Apr 08 02:00:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

