

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066524.D
 Acq On : 7 Apr 2021 22:31
 Operator : JC/MD
 Sample : PB135583ZH#13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB135583ZH#13

Manual Integrations
 APPROVED

MMDadoda
 4/9/2021 12:49:31 PM

Quant Time: Apr 08 02:00:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.587	168	302790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	467194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.350	117	472400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	206577	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	197410	49.51	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.02%		
35) Dibromofluoromethane	7.507	113	159657	50.14	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.28%		
50) Toluene-d8	10.028	98	604453	49.96	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.92%		
62) 4-Bromofluorobenzene	12.345	95	197833	50.42	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.84%		
Target Compounds						Qvalue
16) Acetone	3.736	43	21953	13.13	ug/l	96
18) Methyl Acetate	4.237	43	5511m	1.32	ug/l	
20) Methylene Chloride	4.452	84	16677	4.65	ug/l	92
43) Isopropyl Acetate	8.091	43	30020	3.45	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066524.D
Acq On : 7 Apr 2021 22:31
Operator : JC/MD
Sample : PB135583ZH#13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB135583ZH#13

Manual Integrations
APPROVED

MMDadoda
4/9/2021 12:49:31 PM

Quant Time: Apr 08 02:00:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

