

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066507.D
 Acq On : 7 Apr 2021 14:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 01:53:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	50.000	49.308	1.4	113	0.00
3 P	Chloromethane	50.000	44.984	10.0	112	0.00
4 C	Vinyl Chloride	50.000	45.146	9.7#	108	0.00
5 T	Bromomethane	50.000	43.564	12.9	108	0.00
6 T	Chloroethane	50.000	45.733	8.5	106	0.00
7 T	Trichlorofluoromethane	50.000	45.590	8.8	107	0.00
8 T	Diethyl Ether	50.000	49.530	0.9	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.701	2.6	112	0.00
10 T	Methyl Iodide	50.000	48.302	3.4	106	0.00
11 T	Tert butyl alcohol	250.000	205.940	17.6	105	0.00
12 CM	1,1-Dichloroethene	50.000	46.109	7.8#	110	0.00
13 T	Acrolein	250.000	199.939	20.0	100	0.00
14 T	Allyl chloride	50.000	45.415	9.2	119	0.00
15 T	Acrylonitrile	250.000	220.984	11.6	105	0.00
16 T	Acetone	250.000	205.381	17.8	109	0.00
17 T	Carbon Disulfide	50.000	43.803	12.4	112	0.00
18 T	Methyl Acetate	50.000	41.660	16.7	107	0.00
19 T	Methyl tert-butyl Ether	50.000	49.053	1.9	115	0.00
20 T	Methylene Chloride	50.000	46.206	7.6	113	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.772	6.5	112	0.00
22 T	Diisopropyl ether	50.000	51.056	-2.1	119	0.00
23 T	Vinyl Acetate	250.000	252.571	-1.0	116	0.00
24 P	1,1-Dichloroethane	50.000	48.522	3.0	116	0.00
25 T	2-Butanone	250.000	214.830	14.1	103	0.00
26 T	2,2-Dichloropropane	50.000	47.901	4.2	113	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.186	3.6	114	0.00
28 T	Bromochloromethane	50.000	51.445	-2.9	134	0.00
29 T	Tetrahydrofuran	250.000	212.573	15.0	100	0.00
30 C	Chloroform	50.000	48.441	3.1#	114	0.00
31 T	Cyclohexane	50.000	47.419	5.2	118	0.00
32 T	1,1,1-Trichloroethane	50.000	47.129	5.7	110	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.499	-7.0	133	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	121	0.00
35 S	Dibromofluoromethane	50.000	51.469	-2.9	127	0.00
36 T	1,1-Dichloropropene	50.000	48.834	2.3	117	0.00
37 T	Ethyl Acetate	50.000	42.656	14.7	107	0.00
38 T	Carbon Tetrachloride	50.000	46.806	6.4	109	0.00
39 T	Methylcyclohexane	50.000	50.329	-0.7	118	0.00
40 TM	Benzene	50.000	49.417	1.2	116	0.00
41 T	Methacrylonitrile	50.000	41.368	17.3	102	0.00
42 TM	1,2-Dichloroethane	50.000	49.042	1.9	119	0.00
43 T	Isopropyl Acetate	50.000	45.714	8.6	109	0.00
44 TM	Trichloroethene	50.000	46.415	7.2	110	0.00
45 C	1,2-Dichloropropane	50.000	49.824	0.4#	118	0.00
46 T	Dibromomethane	50.000	48.076	3.8	113	0.00
47 T	Bromodichloromethane	50.000	48.325	3.3	113	0.00
48 T	Methyl methacrylate	50.000	46.687	6.6	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
 Data File : VN066507.D
 Acq On : 7 Apr 2021 14:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 08 01:53:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 06 15:38:34 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	990.134	1.0	114	0.00
50 S	Toluene-d8	50.000	52.485	-5.0	130	0.00
51 T	4-Methyl-2-Pentanone	250.000	224.640	10.1	102	0.00
52 CM	Toluene	50.000	49.903	0.2#	114	0.00
53 T	t-1,3-Dichloropropene	50.000	51.710	-3.4	119	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.119	-0.2	118	0.00
55 T	1,1,2-Trichloroethane	50.000	48.845	2.3	113	0.00
56 T	Ethyl methacrylate	50.000	43.607	12.8	113	0.00
57 T	1,3-Dichloropropane	50.000	49.134	1.7	114	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	246.182	1.5	117	0.00
59 T	2-Hexanone	250.000	194.248	22.3	98	0.00
60 T	Dibromochloromethane	50.000	49.202	1.6	111	0.00
61 T	1,2-Dibromoethane	50.000	48.158	3.7	111	0.00
62 S	4-Bromofluorobenzene	50.000	55.775	-11.5	132	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	121	0.00
64 T	Tetrachloroethene	50.000	46.094	7.8	108	0.00
65 PM	Chlorobenzene	50.000	48.310	3.4	113	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	48.659	2.7	113	0.00
67 C	Ethyl Benzene	50.000	51.294	-2.6#	115	0.00
68 T	m/p-Xylenes	100.000	104.830	-4.8	114	0.00
69 T	o-Xylene	50.000	51.294	-2.6	115	0.00
70 T	Styrene	50.000	45.837	8.3	114	0.00
71 P	Bromoform	50.000	48.908	2.2	109	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	120	0.00
73 T	Isopropylbenzene	50.000	49.583	0.8	114	0.00
74 T	N-amyl acetate	50.000	42.731	14.5	116	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	42.662	14.7	110	0.00
76 T	1,2,3-Trichloropropane	50.000	47.306	5.4	109	0.00
77 T	Bromobenzene	50.000	47.591	4.8	115	0.00
78 T	n-propylbenzene	50.000	51.825	-3.7	116	0.00
79 T	2-Chlorotoluene	50.000	49.528	0.9	116	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.284	-0.6	115	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	43.411	13.2	104	0.00
82 T	4-Chlorotoluene	50.000	51.088	-2.2	116	0.00
83 T	tert-Butylbenzene	50.000	46.280	7.4	114	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.623	-1.2	115	0.00
85 T	sec-Butylbenzene	50.000	49.877	0.2	115	0.00
86 T	p-Isopropyltoluene	50.000	52.121	-4.2	117	0.00
87 T	1,3-Dichlorobenzene	50.000	47.504	5.0	116	0.00
88 T	1,4-Dichlorobenzene	50.000	47.326	5.3	116	0.00
89 T	n-Butylbenzene	50.000	49.390	1.2	123	0.00
90 T	Hexachloroethane	50.000	46.114	7.8	117	0.00
91 T	1,2-Dichlorobenzene	50.000	46.758	6.5	113	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	37.728	24.5	93	0.00
93 T	1,2,4-Trichlorobenzene	50.000	43.463	13.1	112	0.00
94 T	Hexachlorobutadiene	50.000	43.520	13.0	113	0.00
95 T	Naphthalene	50.000	36.193	27.6#	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040721\
Data File : VN066507.D
Acq On : 7 Apr 2021 14:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 08 01:53:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N040621W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 06 15:38:34 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	40.731	18.5	106	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6