

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040722\
 Data File : VN071900.D
 Acq On : 07 Apr 2022 13:26
 Operator : JC\MD
 Sample : N2321-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW7A-GW-20220406

Quant Time: Apr 08 01:13:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	555799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	940266	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	948950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	302277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	299870	47.260	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.520%
35) Dibromofluoromethane	8.022	113	267686	48.381	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.760%
50) Toluene-d8	10.439	98	1186557	52.424	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.840%
62) 4-Bromofluorobenzene	12.727	95	366594	50.080	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.160%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.222	101	27041	5.609	ug/l	95
12) 1,1-Dichloroethene	4.187	96	23942	5.219	ug/l	92
16) Acetone	4.293	43	5216	1.924	ug/l	91
38) Carbon Tetrachloride	8.204	117	7820	1.033	ug/l	90
44) Trichloroethene	9.216	130	606449	96.443	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 08 01:13:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

