

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040723\
 Data File : VN077275.D
 Acq On : 07 Apr 2023 13:13
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.1mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN040723

Quant Time: Apr 07 13:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 07 13:04:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	88	0.00
2 M	Dichlorodifluoromethane	2.852	2.689	5.7	90	0.00
3 M	Chloromethane	3.789	3.678	2.9	92	0.00
4 M	Vinyl Chloride	5.427	5.365	1.1	94	0.00
5 M	Bromomethane	4.090	3.899	4.7	99	0.00
6 M	Chloroethane	3.638	3.771	-3.7	97	0.00
7 M	Trichlorofluoromethane	5.053	4.841	4.2	92	0.00
8 T	Diethyl Ether	2.154	2.152	0.1	94	0.00
9	1,1,2-Trichlorotrifluoroeth	3.160	3.168	-0.3	94	0.00
10 M	1,1-Dichloroethene	3.161	3.002	5.0	93	0.00
11	Methyl Iodide	3.571	3.339	6.5	86	0.00
12	Methyl Acetate	4.710	4.423	6.1	89	0.00
13 M	Acrolein	0.585	0.568	2.9	88	0.00
14 M	Acrylonitrile	1.610	1.507	6.4	90	0.00
15 M	Acetone	0.584	0.535	8.4	74	0.00
16 M	Carbon Disulfide	8.344	8.097	3.0	94	0.00
17	Allyl chloride	4.069	3.840	5.6	85	0.00
18 M	Methylene Chloride	3.671	3.556	3.1	94	0.00
19 M	trans-1,2-Dichloroethene	3.437	3.304	3.9	93	0.00
20 T	Diisopropyl ether	9.699	9.169	5.5	91	0.00
21 M	1,1-Dichloroethane	6.147	6.047	1.6	94	0.00
22 M	cis-1,2-Dichloroethene	4.190	4.112	1.9	96	0.00
23 M	tert-Butyl Alcohol	0.716	0.670	6.4	84	0.00
24 M	Methyl tert-Butyl Ether	11.381	10.971	3.6	91	0.00
25 M	Chloroform	6.662	6.479	2.7	94	0.00
26	Cyclohexane	5.342	5.190	2.8	94	0.00
27 s	1,2-Dichloroethane-d4	2.544	2.654	-4.3	94	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
29	1,1-Dichloropropene	0.650	0.628	3.4	96	0.00
30 M	2-Butanone	0.307	0.276	10.1	83	0.00
31	2,2-Dichloropropane	0.798	0.751	5.9	94	0.00
32 M	1,1,1-Trichloroethane	0.800	0.738	7.8	90	0.00
33 M	Carbon Tetrachloride	0.683	0.631	7.6	90	0.00
34 M	Benzene	2.011	1.926	4.2	94	0.00
35	Methacrylonitrile	0.290	0.285	1.7	98	0.00
36 M	1,2-Dichloroethane	0.630	0.613	2.7	94	0.00
37 M	Trichloroethene	0.492	0.467	5.1	95	0.00
38	Methylcyclohexane	0.881	0.831	5.7	94	0.00
39 M	1,2-Dichloropropane	0.499	0.474	5.0	93	0.00
40	Dibromomethane	0.360	0.336	6.7	92	0.00
41 M	Bromodichloromethane	0.717	0.677	5.6	92	0.00
42 M	Vinyl Acetate	0.889	0.814	8.4	90	0.00
43	Ethyl Acetate	0.569	0.514	9.7	89	0.00
44	Isopropyl Acetate	1.018	0.954	6.3	90	0.00
45 T	1,4-Dioxane	0.011	0.010	9.1	89	0.00
46	Methyl methacrylate	0.438	0.408	6.8	92	0.00
47	n-amyl Acetate	0.890	0.853	4.2	95	0.00
48 M	t-1,3-Dichloropropene	0.803	0.739	8.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN040723\
 Data File : VN077275.D
 Acq On : 07 Apr 2023 13:13
 Operator : JC\MD
 Sample : VSTDICV020
 Misc : 5.1mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN040723

Quant Time: Apr 07 13:30:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N040723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Apr 07 13:04:49 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	cis-1,3-Dichloropropene	0.854	0.805	5.7	93	0.00
50 M	1,1,2-Trichloroethane	0.522	0.490	6.1	94	0.00
51	Ethyl methacrylate	0.802	0.769	4.1	95	0.00
52	1,3-Dichloropropane	0.872	0.805	7.7	92	0.00
53 M	Dibromochloromethane	0.551	0.510	7.4	92	0.00
54 M	1,2-Dibromoethane	0.524	0.494	5.7	92	0.00
55 M	2-Chloroethyl vinyl ether	0.248	0.210	15.3	83	0.00
56 M	Bromoform	0.370	0.333	10.0	89	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
58 M	4-Methyl-2-Pentanone	0.562	0.528	6.0	92	0.00
59 M	2-Hexanone	0.434	0.399	8.1	89	0.00
60 S	4-Bromofluorobenzene	0.519	0.516	0.6	96	0.00
61 M	Tetrachloroethene	0.357	0.334	6.4	97	0.00
62 M	Toluene	2.093	1.983	5.3	96	0.00
63 S	Toluene-d8	1.055	1.056	-0.1	96	0.00
64 M	Chlorobenzene	1.272	1.196	6.0	96	0.00
65	1,1,1,2-Tetrachloroethane	0.468	0.438	6.4	93	0.00
66 M	Ethyl Benzene	2.319	2.177	6.1	95	0.00
67 M	m/p-Xylenes	0.909	0.844	7.2	96	0.00
68 M	o-Xylene	0.909	0.866	4.7	97	0.00
69 M	Styrene	1.457	1.371	5.9	96	0.00
70	Isopropylbenzene	2.308	2.207	4.4	97	0.00
71 M	1,1,2,2-Tetrachloroethane	0.707	0.664	6.1	95	0.00
72	1,2,3-Trichloropropane	0.480	0.497	-3.5	98	0.00
73	Bromobenzene	0.469	0.460	1.9	99	0.00
74	n-propylbenzene	2.553	2.463	3.5	99	0.00
75	2-Chlorotoluene	1.530	1.456	4.8	96	0.00
76	1,3,5-Trimethylbenzene	1.878	1.796	4.4	96	0.00
77	t-1,4-Dichloro-2-butene	0.245	0.224	8.6	93	0.00
78	4-Chlorotoluene	1.445	1.404	2.8	100	0.00
79	tert-butylbenzene	1.635	1.561	4.5	95	0.00
80	1,2,4-Trimethylbenzene	1.831	1.748	4.5	95	0.00
81	sec-Butylbenzene	2.335	2.224	4.8	95	0.00
82	p-Isopropyltoluene	1.870	1.776	5.0	95	0.00
83 M	1,3-Dichlorobenzene	0.889	0.832	6.4	98	0.00
84 M	1,4-Dichlorobenzene	0.862	0.812	5.8	98	0.00
85	n-Butylbenzene	1.546	1.458	5.7	99	0.00
86 T	Hexachloroethane	0.398	0.357	10.3	92	0.00
87 M	1,2-Dichlorobenzene	0.871	0.825	5.3	95	0.00
88	1,2-Dibromo-3-Chloropropane	0.135	0.118	12.6	92	0.00
89	1,2,4-Trichlorobenzene	0.343	0.296	13.7	99	0.00
90	Hexachlorobutadiene	0.176	0.165	6.2	98	0.00
91 M	Naphthalene	1.309	1.097	16.2	93	0.00
92	1,2,3-Trichlorobenzene	0.334	0.278	16.8	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0